

DEPARTMENT OF CHEMISTRY
&
CENTRE OFF ADVANCE STUDIES IN CHEMISTRY
PANJAB UNIVERSITY CHANDIGARH –160014 (INDIA)

No.CHM/16/.....

Dated:

Please send the quotations in two bid shape so as to reach undersigned on 16.11.2016.

- **Technical bid consisting of all technical details alongwith commercial terms and conditions.**
- **Financial bid indication items-wise price for the items mentioned in the technical bid.**
- **The technical bid and the financial bid should be sealed by the bidder in separate cover duly super scribed and both these sealed covers are to be put in a bigger cover which should also be sealed and duly super scribed. To reach on or before**
- **And also send 2% cost of the Instrument as EMD quoted on the Quotation if the amount exceed Rs.50,000/-.**

Incubator =1

Specifications:

Incubators for all gentle incubation applications; outstanding dynamics keep the required temperature virtually stable with homogenous distribution; irrespective of how many time the door is opened. For applications that have to meet high standards of thermal safety, the excellent class excellent class is recommended, extensively equipped for documentation. A multifunctional, PID process controller ensures absolutely reliable temperature control and the highest process safety.

Process Control

Throttle valve for smooth adjustment of air changes inside the incubator. All of our incubators are available with forced air circulation for superior temperature homogeneity and fastest recovery times. User friendly, compact in size and excellent value for money. Our models are easy to clean and decontaminate, with corrosion-resistant stainless steel interiors and removable shelves.

Performance Features

- PID Controller with integrated time 0 to 99 hours, one ramp and soak function
- Engineered to provide at even temperature distribution and rapid heating
- Digital temperature setting with an accuracy of 0.1° C.
- Adjustable ventilation by means of top exhaust duct.
- Safety devices to prevent temperature overshoots
- Alarms for temperature deviation
- Chrome plated stainless steel racks
- Rounded edges for easy cleaning
- Door with glass window

1. The quality, make, packing and other particulars of each item should clearly be stated/indicated in your quotation.
2. It should clearly be stated whether the rates are inclusive / exclusive of SALE TAX etc.
3. The rates of INSURANCE if any should be specified. The firm will be required to submit original payees receipt alongwith bill. NO PAYMENT OF INSURANCE CHARGES IS PERMISSIBLE UNLESS THESE REQUIREMENTS ARE COPLIED WITH.
4. THE QUOTATION SHOULD ALWAYS BE SUBMITTED IN DUPLICATE.
5. Please quote only for the items available in 'READY STOCK.'
6. Please quote only for the make required and strictly according to the specifications etc.
7. Please also indicate the DISCOUNT separately, if any.
8. While submitting the quotation please write on the OUTER ENVELOPE, '**QUOTATIONS TO BE OPENED ON 17.11.2016**' and the envelope be put in another envelope address to **Dr. Gurpreet kaur, Investigator Incharge Scheme, Department of Chemistry, Panjab University, Chandigarh – 160014**
9. Please send the quotation by post along with manufacturer/distributor certificate .